

# Notice of Allowability

Application No.

10/085,070

Examiner

Brian Ensey

Applicant(s)

GARBER ET AL.

Art Unit

2615

## -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to after final amendment dated 8/18/06.
2. ☒ The allowed claim(s) is/are 4, 6-12, 14, 16, 18, 19, 31 and 33 renumbered 1-14.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☐ All b) ☐ Some\* c) ☐ None of the:
    1. ☐ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.
  - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached
    - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
  - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

## Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/08), Paper No./Mail Date \_\_\_\_\_
4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☐ Interview Summary (PTO-413), Paper No./Mail Date \_\_\_\_\_
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other \_\_\_\_\_

## **DETAILED ACTION**

### ***Allowable Subject Matter***

Claims 4, 6-12, 14, 16, 18, 19, 31 and 33 are allowed.

The following is an examiner's statement of reasons for allowance: The present invention is directed to a hearing aid storage case with activity detection. Independent claims 4 and 16 identify the uniquely distinct feature of a first visual cue indicative that electrical signals produced by a detection transducer in the case that the hearing aid is oscillating and a second visual cue not indicative that electrical signals produced by a detection transducer in the case that the hearing aid is oscillating in combination with all the limitations as disclosed in claims 4 and 16. Independent claim 12 identifies the uniquely distinct feature of a first visual cue activated when electrical signals produced by a detection transducer in the case are above a predetermined threshold level and a second visual cue activated when electrical signals produced by a detection transducer in the case are below a predetermined threshold level in combination with all the limitations as disclosed in claim 12. Independent claim 31 identifies the uniquely distinct feature of a switch mechanism activated by the case lid to power the logic circuit for a preset time after the lid is closed to determine that electrical signals produced by a detection transducer in the case that the hearing aid is oscillating and activating a first visual cue to signify if the hearing aid is properly turned off or left on in combination with all the limitations as disclosed in claim 31. Independent claim 33 identifies the uniquely distinct feature of monitoring a noise generated by a turned on hearing aid and storing the noise in a memory as a reference signal to determine if the electric signals are indicative of the hearing aid being turned off in combination with all the

Art Unit: 2615

limitations as disclosed in claim 33. The closest prior art, Power et al. (US 4,095,057) teaches a frequency response testing apparatus for hearing aids with visual indicators of operation and Mattatall (US 4,379,988) teaches a hearing aid storage case and base. The prior art fails to anticipate or render the independent claims obvious.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian Ensey whose telephone number is 571-272-7496. The examiner can normally be reached on Monday - Friday 6:30 AM - 3:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Sinh Tran can be reached on 571-272-7564. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

**Any response to this action should be mailed to:**

Commissioner of Patents and Trademarks  
P.O. Box 1450  
Alexandria, Va. 22313-1450

**Or faxed to:**


(571) 273-8300, for formal communications intended for entry and for informal or draft communications, please label "PROPOSED" or "DRAFT".  
Hand-delivered responses should be brought to:

Customer Service Window  
Randolph Building  
401 Dulany Street  
Arlington, VA 22314

Art Unit: 2615

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

BKE  
August 22, 2006

  
VIVIAN CHIN  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 2600

8/22/06